AMDA-Nepal

Siddhartha Children And Women Hospital

Butwal-7, Rupandehi

Technical Specification Operating table with Ortho attachment

S.N.	Purchaser's Specifications	Bidder's Compliance (Yes/No)	Reference Page No.	Remarks
	O.T Table five to seven functional.			
	Manufacturer:			
	Brand:			
	Type / Model:			
	Country of Origin:			
1	Description of Function:			
2	The table is desired suitable to be used for generally surgery, Obs/Gyn, neck, thorax, abdomen, perineum and extremities surgery as well as ENT and orthopedic operation.			
3	Two control panels including remote panel and keypad panel with battery backup. There are also electric lumbar bridges and manual/electric brakes.			
4	There are multiple function like up and down, right tilt position, left tilt position, anorectic operation position, trendelenburg position, reverse trendelenburg, lateral arm support, gyne position, ENT operation.			
5	Length and width for table top (2070*525mm) elevation (up/down) 680- 980mm with load capacity 270with 70mm thick memory sponge mattress.			
6	Main body materials and all parts of accessories is manufacture from stainless still.			
7	Accessories, Spares and Consumable:			
7.1	All stander accessories, consumables and parts required to operate the equipment .Including all stander tolls .Bidders must specify the quantity of every item include in their offer (including items not specified above)			
8	Installation and Commissioning & User Training			
8.1	Supplier must accomplish proper installation and user training in site.			
9	Warranty & Maintenance Service During Warranty Period			
9.1	Comprehensive warranty of all accessories up to two years.			
10	Authorization			
10.1	Manufacturer's Authorization or local Distributor Authorization (Manufacturer's Authorization to the main Distributor is also required in case of Local Authorization).			
11	Documentation			
11.1	 User (Operating) manual in English. Service (Technical/Maintenance) manual in English. List of important spare parts and accessories with their numbers and costing. Certificate of calibration and inspection from factory. Shall accomplish all the qualification criteria of the bidder mentioned in the main bid document. 			