

Sealed Quotation Form for the Supply of

“Autoclave Machine”

In

AMDA Mechi Hospital, Mechinagar 10, Dhulabari, Jhapa
(ID no.:0105/AMH/105/2081/082)

“Technical & Financial Proposal”



AMDA Mechi Hospital
Mechinagar 10, Dhulabari, Jhapa
Phone No.:023-564550
E-mail: amh.dhulabari@amda.org.np
URL: www.amda.org.np



SUPPLIER'S PROFILE



Name of Company / Firm:

Company/Firm Registration No.:

Address:

VAT/PAN No.:

Validity/Expiry of VAT/Pan Certificate:

Phone No.: Fax No.:

E-mail:

Web site:

Type of Transaction: Retailer/Stockiest/Supplier/Manufacturer

Others (Specify).....

Mailing Address:

.....
.....
.....
.....

I/We confirm the accuracy of the information given in the sealed quotation document and fully responsible to supply the items as cited under the following terms & condition.

Signature:

Stamp:

Name of Proprietor:

Date:

AMDA Mechi Hospital
Mechinagar 10, Dhulabari, Jhapa

Notice of sealed quotation/proposal invitation
(Notice No.: 0105/AMH/105/2081/82)



Bidder's Qualification checklist

SN	Description	Enclosure	
		Yes	No
A	Mandatory		
1	Cover letter for the submission of quotation/proposal documents		
2	Company/firm registration/renewal certificate		
3	PAN/VAT registration certificate		
4	Tax clearance certificate (F/Y 2080/81)		
5	Self-declaration letter		
6	Declaration of acceptance of terms and condition of bid document specified in detail notice, bidder's qualification bid form, technical bid form and financial bid form.		
7	Signature of bidder's representative in each page of document and firm/company stamp		
8	Authorization Letter		

Name of firm/company:

Representative:

Contact number:

Signature:

Firm/Company stamp:

आम्दा मेची अस्पताल, धुलावारी झापाको
दरभाउपत्र आव्हानको सूचना

(प्रथम पटक प्रकाशित मिति: २०८१।१०।०२)

आम्दा मेची अस्पताल, धुलावारी झापाको प्रयोजनका लागि मेडिकल उपकरण (Autoclave Machine) खरिद गर्नुपर्ने भएकोले ईजाजत प्राप्त ईच्छुक आपूर्तिकर्ता/कम्पनिहरुबाट यो सूचना (सूचना नं. 0105/AMH/105/2081/082) प्रकाशित भएको मितिले ७ दिन अर्थात मिति २०८१।१०।०८ गते भित्र आम्दा मेची अस्पताल, धुलावारी आइपुग्ने गरि निम्न शर्तहरुको अधिनमा रहि संलग्न ढाँचामा दरभाउ पत्र प्रस्ताव आह्वान गरिएको छ ।
शर्तहरु:

१. ईच्छुक बोलपत्र/दरभाउपत्र खरिदकर्ताले दरभाउपत्र आम्दा मेची अस्पतालमा दरभाउ पत्र पेश गर्नु पर्नेछ ।
२. शिलबन्दी दरभाउपत्र खरिदकर्ताले बोलपत्र/दरभाउपत्र फारम निशुल्क वेबसाईटबाट प्राप्त गर्न सकिनेछ ।
३. शिलबन्दी दरभाउपत्र मिति २०८१।१०।०८ गते २ बजेभित्र पेश गर्नुपर्नेछ ।
४. शिलबन्दी दरभाउपत्र मिति २०८१।१०।०८ गते दिनको ३ बजे आम्दा मेची अस्पताल, धुलावारीमा बोलपत्रदाता वा निजको प्रतिनिधिको रोहवरमा खोलिनेछ । बोलपत्रदाता वा निजको प्रतिनिधि उपस्थित नभएमा पनि उक्त बोलपत्र/दरभाउपत्र खोल्न बाधा पुग्ने छैन । साथै उक्त दिन सार्वजनिक विदा परेमा विदा पछि कार्यालय खुलेको दिन तोकिएको स्थान र समयमा बोलपत्र/दरभाउपत्र खोलिने छ ।
५. आपूर्तिकर्ता/प्रस्तावकले प्रस्तावको साथमा Autoclave Machine को पूर्ण रुपमा सञ्चालन गर्न आवश्यक पर्ने Technical Specification मा समावेश भए बमोजिमका सम्पूर्ण सामग्रीहरु सहितको शिलबन्दी दरभाउ प्रस्ताव पेश गर्नुपर्नेछ ।
६. आपूर्तिकर्ताले प्रस्ताव गरेको Autoclave Machine को प्राविधिक विवरणमा उल्लेख भएका सम्पूर्ण कागजातहरु पेश गर्नुपर्नेछ ।
७. आपूर्तिकर्ताले प्रस्ताव गर्ने Autoclave Machine विग्रेको वा काम नगरेको अवस्थामा कम्तिमा एक वर्षसम्म निशुल्क मर्मत संभार र तोकिएको समयावधीभित्र मर्मतसंभार हुन नसकेको खण्डमा नयाँ मेशिन सट्टा गर्नुपर्नेछ ।
८. आपूर्तिकर्ताले प्रस्ताव गर्ने Autoclave Machine को प्रयोगकर्तालाई निशुल्क तालिम दिनुपर्नेछ ।
९. छनौट भएको बोलपत्र/दरभाउपत्रदाता अर्थात् आपूर्तिकर्ता र आम्दा मेची अस्पताल बिच संझौता सम्पन्न भइ खरिद आदेश जारी गरेकोमा आपूर्तिकर्ताले खरिद आदेश प्राप्त गरेको मितिले बढिमा ३० दिन भित्र आम्दा मेची अस्पतालमा संझौता बमोजिमको Autoclave Machine जडान तथा परिक्षण गरि सकुपर्नेछ ।
१०. प्रस्ताव छनौट गर्ने नगर्ने पूर्ण अधिकार आम्दा नेपालमा सुरक्षित रहने छ । प्रस्ताव छनौट भएको प्रस्तावक/ आपूर्तिकर्ताले छनौटको सूचना पाएको मितिले ७ दिन भित्र संझौता गरिसकुपर्ने छ । तोकिएको समयसिमा भित्र संझौता गर्न नआएमा प्रस्ताव छनौटको सूचना स्वतः रद्द हुनेछ र संबन्धित प्रस्तावकलाई कालोसुचीमा राख्न सकिनेछ ।
११. छनौट भएको आपूर्तिकर्ताले खरिद आदेशमा तोकिएको समय र स्थानमा सामान आपूर्ति गर्न असमर्थ भएमा संबन्धित आपूर्तिकर्तालाई कालो सुचीमा समेत राख्न सकिनेछ ।
१२. बोलपत्रका साथ अध्यावधिक कम्पनी/फर्म दर्ता प्रमाणपत्रको प्रतिलिपि, ब्यबसाय संचालनगर्न छुट्टै स्विकृत लिनु पर्नेमा स्विकृत लिएको / नविकरण खुलेको प्रमाणपत्रको प्रतिलिपि, अघिल्लो आ/व को करचुक्ताको प्रमाणपत्र, कर दर्ता प्रमाणपत्रको प्रतिलिपि संलग्न गर्नुपर्नेछ ।
१३. रितपूर्वक प्राप्त नभएको दरभाउपत्र उपर कुनै कारवाही हुने छैन ।
१४. खरिद संझौता १ वर्ष सम्मको लागि मान्य हुने गरि गरिनेछ । संझौता अवधिभर दरभाउपत्रमा उल्लेखित प्रस्तावको दर तथा विवरण परिवर्तन मान्य हुने छैन ।
१५. आम्दा मेची अस्पताल, धुलावारीलाई आपूर्ति गर्ने सरसामानहरु उत्पादकले प्याकिङ्ग गरि बजारमा वितरण गर्दाको अवस्थामा हुनुपर्नेछ । क्षतिग्रस्त सरसामान स्विकार गरिने छैन । आम्दा मेची अस्पताल, धुलावारीले आपूर्तिकर्ताबाट प्राप्त Autoclave Machine को परिक्षण प्रतिवेदन पास नगर्दासम्म सरसामानहरुको क्षति हुन पुगेमा सो को जिम्मेवारी आपूर्तिकर्ताले लिनुपर्नेछ ।



१६. आम्दा मेची अस्पताल, धुलावारीले तोकेको समयभित्र तोकिएका Autoclave Machine आपूर्तिकर्ताले बुझाउन नसकेमा आम्दा मेची अस्पतालले अन्य पक्षसँग उक्त Autoclave Machine खरिद गर्न सकेछ । आम्दा मेची अस्पतालले अन्य आपूर्तिकर्ताबाट खरिद गरेको सरसामानहरूको सम्पूर्ण मूल्यको जरिवाना वा बैंक ग्यारेन्टी (परफरमेन्स वण्ड) जफत गरि आपूर्तिकर्तासँग भएको संझौता भङ्ग गर्न सकेछ ।
१७. आम्दा मेची अस्पताल, धुलावारीले जारी गरेको खरिद आदेश बमोजिम सम्पूर्ण बस्तु, सेवा तथा सरसामानहरू तोकिएको स्थानमा बुझाई सके पश्चात् समेत सो को परिक्षण पछि मात्र नेपाली मुद्रामा क्रस चेक मार्फत भुक्तानी गरिनेछ । नेपाल सरकारको प्रचलित नियमानुसार लाग्ने कर अग्रिम कट्टा गरिनेछ ।
१८. संझौता सम्पन्न गर्दा तोकिएका बस्तु, सेवा वा सरसामानहरूको गुणस्तरको सुनिश्चितता, आवश्यक प्राविधिक सहयोग र अन्य सबै प्रकारका सरसहयोग, दरभाउपत्र/बोलपत्र साथ पेश गरे बमोजिम आपूर्तिकर्ताले उपलब्ध गराउनु पर्नेछ ।
१९. शिलबन्दि दरभाउपत्र पेश गर्दा जम्मा रकम नेपाली मुद्राको दररेट अंक र अक्षर दुबैमा स्पष्ट बुझिने गरी भर्नुपर्नेछ । अंक र अक्षरमा दुविधा भएमा अक्षरमा लेखिएको दर मान्य हुनेछ । केरमेट भएको ठाउँमा सहीछाप गर्नु पर्नेछ । बोलपत्र/दरभाउपत्रको प्रत्येक पानामा बोलपत्रदाताले सहीछाप गरेको हुनुपर्नेछ । त्यसो नभएको पाईएमा बोलपत्र फारम रद्द गर्न सकिनेछ । नेपाल सरकारलाई तिर्न बुझाउन पर्ने भ्याट बाहेक सबै प्रकारका करहरू र स्थानीय स्तरमा तिर्नुपर्ने अन्य सबै प्रकारका कर तथा शुल्कहरूको दररेटमा समावेश भएको मानिनेछ । बोलपत्रदाताले तोकिएको वस्तु तथा सेवाको आपूर्ति, मर्मत गर्नुपर्ने स्थान, अवधि, शर्तहरू आदिको पूर्ण विवरण बुझी जानी दररेट भरेको मानिनेछ । प्राप्त भएका प्रस्तावहरू मध्ये प्राविधिक अध्ययन/मुल्यांकन एवम आर्थिक अध्ययन/मुल्यांकनका आधारमा उपयुक्त प्रस्तावक संग सम्झौता गरिने छ ।
२०. शिलबन्दी दरभाउपत्रमा संसोधन आवश्यक देखिएमा पेश गर्ने अन्तिम मिति अगावै आम्दा मेची अस्पतालले संसोधन गर्न सकेछ । अन्य शर्त तथा कार्य विवरण बोलपत्र/दरभाउपत्रमा उल्लेख भए बमोजिम हुनेछ । दरभाउपत्र/बोलपत्रहरू आंशिक वा पूर्णरूपमा, जानकारी दिई वा नदिई स्विकृत गर्ने वा नगर्ने सम्पूर्ण अधिकार, पूर्ण रूपमा आम्दा मेची अस्पतालमा सुरक्षित रहनेछ ।

शिलबन्दि दरभाउपत्र संबन्धि थप जानकारीको लागि तलको ठेगानामा सम्पर्क गर्न सकिनेछ ।

सम्पर्क राख्ने ठेगाना

आम्दा मेची अस्पताल धुलावारी, झापा

फोन: ०२३ ५६४५५०

इमेल: amh.dhulabari@amda.org.np

आम्दा मेची अस्पताल, धुलावारीको उपरोक्त सूचना बमोजिम तोकिएका उपरोक्त शर्तहरू पूर्णरूपमा स्विकार गर्दछु / गर्दछौं ।

संस्थापक वा आधिकारीक पदाधिकारीको नाम:

हस्ताक्षर:

मिति:

फर्म/कंपनीको छाप:



AMDA Mechi Hospital, Dhulabari, Jhapa
(ID no.:0105/AMH/105/2081 82)

Technical Proposal for Autoclave Machine

S.NO	Purchase Specifications	Bidders Compliance Sheet		
		YES/ NO	Page no in Catalogue	Remarks
Autoclave Machine				
Manufacturer :-				
Brand :-				
Type / Model :-				
Country of Origin :-				
1	Description of Function			
1.1	Autoclave is required for sterilizing objects in high temperature and high-pressure steam.			
2	Operational Requirements			
2.1	Autoclave must be user friendly interfaces allows the user to use very easily and can sterilize wrapped/unwrapped, non-porous and solid instrument.			
3	System Configuration			
3.1	Autoclave made of interior and exterior at least 304 Ti grade stainless steel and all the pipes and fittings should be constructed of stainless steel and brass with complete accessories			
4	Technical Specifications			
4.1	The autoclave should provide sterilization minimum at 121°C and 150°C for both wrapped and unwrapped tools and also flash cycle for rapid sterilization.			
4.2	The autoclave should be equipped with a powerful vaccum pump to eject air pocket from the chamber at the beginning and at the end of cycle (pre-vacuum, post vacuum and chamber 100%)			
4.3	It should have minimum four sterilization programs and at least two or more test program and regulates sterilization time and activates alarm when sterilization cycle completes			
4.4	Minimum volume at least 250 liters having working pressure at least 5 psi to 34 psi.			
4.5	Should be fully automatic cylindrical horizontal autoclaves with tripled walled construction, front-loading with PLC/HMI automation hardware based and in horizontal shapes with triple layer with boiler autoclave and also have User-friendly Control Panel			
4.6	It should be at least class B autoclave so that hollow bodied instrument, hand pieces and turbines can be fully autoclaved having minimum 7" (inches) touch screen and the information shows in display like cycle stage, chamber temperature, chamber pressure along with the information about failures and interrupts.			
4.7	It should be thermal printer with last 100 cycles storage capacity or more and having aluminium for insulation.			
4.8	Dimension of inner size: At least (500 * 1200) mm or more and having Loading Capacity 12 KW or less.			

AMDA-Nepal



4.9	It should be Silicon jointless rubber door gasket and automatic pressure control switch and should be digital temperature controller with timer and alarm .			
5	Accessories, spares and consumables			
5.1	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above).			
6	Operating Environment			
6.1	The system offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power Supply, Climate, Temperature, Humidity, etc.			
6.2	Power supply: At least 220-240V AC, 50Hz fitted with appropriate plug type D round 3 pins. The power cable must be at least 3 meter in length.			
7	Standards and Safety Requirements			
7.1	Must submit ISO13485:2003/AC:2007 for Medical Devices			
7.2	Minimum CE (93/42 EEC Directives) or USFDA approved product certificate.			
8	User Training			
8.1	Must provide user training (including how to use and maintain the equipment).			
9	Warranty			
9.1	Comprehensive warranty for 1 year after acceptance and bring into complete operation.			
10	Maintenance Service During Warranty Period			
10.1	During the warranty period supplier must ensure corrective/breakdown maintenance whenever required.			
11	Installation and Commissioning			
11.1	Supplier must accomplish proper installation and commissioning of the equipment on site.			
12	Documentation			
12.1	The bidder should compulsorily fill the technical specification tender form and clearly mention the Manufacturer , Brand, Model and Country of Origin.			
12.2	The bidder should submit the original brochure or e-copy.			
12.3	User (Operating) and Service (Technical / Maintenance) manual in English.			
12.4	Bidder must completely fill the Technical Specification Form (TSF). Only Yes/no/all complies should not be written. Page number in the catalogue of all the required parameters must be clearly mentioned and highlighted. Failure in doing so may lead to rejection of bid from technical committee. Yes ,Comply			



 [Signature]

 08/11/021



AMDA Mechi Hospital, Dhulabari, Jhapa
Financial Proposal for Autoclave Machine (0105/AMH/105/2081/82)

Bidders/Suppliers:

Name:
Address:
Contact Number :
PAN/VAT No :

Sn.	Description	Specification	Unit	Quantity (A)	Estimated Price (Inclusive of all government tax)	Unit Cost (B)	VAT (if applicable) (C)	Total Cost [D=B+C]	Remarks
1	Equipment:	Autoclave Machine	Set	1	1,000,000.00				
2	Capacity/Specification:	250L (Fully Automatic)							
3	Standard:	Must submit ISO13485:2003/AC:2007 for Medical Devices AND CE (93/42 EEC Directives) or USFDA approved product certificate.							
4	Brand:	Must specify.							
5	Model:	Must specify.							
Total Amount									
Total Amount in Words:									

Submitted By:
Signature:
Date:
Stamp: